

Technical Note

Filling QuantIFERON®-TB Gold blood collection tubes

The QuantIFERON-TB Gold assay (QFT®) measures the Interferon-gamma (IFN- γ) response in whole blood stimulated with antigen. The kit uses specialized QFT blood collection tubes. The following is a guide for blood collection into these tubes. (Catalogue No: T0590-0301).

FILLING TUBES

One QFT test uses the following three collection tubes:

- Nil Control (Grey cap)
- TB Antigen (Red cap)
- Mitogen Control (Purple cap)



0.8 mL

1.0 mL

1.2 mL

Approved fill volume range; image for guidance only.

These procedures should be followed for optimal results:

- Tubes should be at 17 - 25°C at the time of blood filling.
- Collect 1 mL of blood by venipuncture directly into each QFT blood collection tube in the order Nil, TB-Antigen and Mitogen.

As 1 mL tubes draw blood relatively slowly, keep the tube on the needle for 2-3 seconds once the tube appears to have completed filling to ensure that the correct volume is drawn.

- The black mark on the side of the tubes indicates the 1 mL fill volume. QFT blood collection tubes have been validated for volumes ranging from 0.8 to 1.2 mL. If the level of blood in any tube is not close to the indicator line, it is recommended to obtain another blood sample. As a guide, the picture on the left illustrates the approved fill range.

If a "butterfly needle" is used, prime tubing with a "purge" tube before filling the QFT tubes.

MIXING TUBES

Antigens have been dried onto the inner wall of the blood collection tubes. It is essential that the tubes' contents be thoroughly mixed with the blood. Thorough mixing will dissolve the heparin, preventing clotting, and allow resolubilization of the stimulating antigen. Mixing is performed by shaking the tubes ten (10) times just firmly enough to ensure that the entire inner surface of the tube is coated with blood.

- The tubes must be transferred to a $37^{\circ}\text{C} \pm 1^{\circ}\text{C}$ incubator as soon as possible, and within 16 hours of collection. Prior to incubation, maintain tubes at room temperature ($22^{\circ}\text{C} \pm 5^{\circ}\text{C}$) Do not refrigerate or freeze the blood samples. Tubes should be incubated **upright**.

If the blood is not incubated immediately after collection, mixing of the tubes by inverting 10 times must be performed immediately prior to incubation.

- Over-energetic shaking may cause gel disruption and could lead to aberrant results.

Further information on QFT tube handling can be found at www.cellestis.com



A correctly mixed tube.

Full instructions for use can be found in the Package Insert, available in up to 25 different languages, at www.cellestis.com

For more information, please contact the Cellestis office nearest you or visit www.cellestis.com.

World Headquarters

Cellestis Limited
Email: info@cellestis.com
Tel: +61 3 8527 3500

North America / South America

Cellestis Inc.
Email: customer.service@cellestis.com
Tel: +1 661 775 7480 (outside USA)
Toll free: 800 519 4627 (USA only)

Europe / Middle East / Africa

Cellestis GmbH
Email: europe@cellestis.com
Tel: +49 6151 428 59 0

Asia / Pacific

Cellestis AP Pte Ltd
Email: asiapac@cellestis.com
Tel: +65 6322 0822

Australia / New Zealand

Cellestis International
Email: anz@cellestis.com
Tel: +61 3 8527 3500

Japan / Korea

Cellestis Asia KK
Email: jp.kr@cellestis.com